

ASTM E2779 Pellet Heater Run Sheets

Client: Thelin Job Number: 19-501 Tracking #: 0038
Model: Providence Run Number: 1 Test Date: 9/10/2019

Pellet Heater Control Settings

High Burn Rate Settings: Control on "High," Fan Trim max

Medium Burn Rate Settings: Control on "Low," Fan Trim 8.0V

Low Burn Rate Settings: Control on "Low," Fan Trim 7.0V

Preburn Notes

Preburn Start Time: 8:30

Time	Notes

Test Notes

Test Burn Start Time: 9:30

Time	Notes
60:00	Changed Filter A Changed to Med settings
180:00	
360:00	

Test Burn End Time: 15:30

Flue Gas Concentration Measurement


Calibration Gas Values: Span Gas CO₂ (%): 16.93 CO (%): 4.33
Mid Gas CO₂ (%): 10.00 CO (%): 2.51

Calibration Results:

	Pre Test			Post Test		
	Zero	Mid	Span	Zero	Mid	Span
Time	8:39	8:42	8:40	16:20	16:22	16:23
CO ₂	0.00	10.05	17.01	-0.01	10.03	16.96
CO	0.000	2.480	4.311	-0.003	2.466	4.259

Flue Gas Probe Leak Check: Initial: No Leakage

Final: No Leakage

Technician Signature: 

Date: 9/26/2019

ASTM E2515 - Glass Filters

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
3577	121.4	121.5	-	-	SB		
3578	116.1	115.9	-	-	SB	19-510	1
3579	123.6	123.2	123.4	-	SB		
3580	122.4	122.6	-	-	SB		
3581	116.7	116.8	-	-	SB		
3582	116.3	116.2	-	-	SB		
3583	124.1	124.1	-	-	SB	19-511	1
3584	116.3	116.4	-	-	SB		
3585	121.4	121.4	-	-	SB		
3586	124.1	124.1	-	-	SB		
3587	116.4	116.4	-	-	SB		
3588	115.8	115.9	-	-	SB	19-501	1
3589	123.3	123.5	-	-	SB		
3590	121.6	122.0	121.9	-	SB		
3591	115.4	115.5	-	-	SB		
3592	116.2	116.0	-	-	SB		
3593	123.8	123.9	-	-	SB	19-517	1
3594	121.6	121.7	-	-	SB		

Weight 1 Date/Time:

7/24-8:00

Weight 2 Date/Time:

7/25-7:00

Weight 3 Date/Time:

7/26-7:45

Weight 4 Date/Time:

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
3595	116.2	116.0	-	-	SB	19-517	1
3596	115.3	115.5	-	-	SB		
3597	123.6	123.8	-	-	SB		
3598	116.0	115.9	-	-	SB		
3599	121.8	121.8	-	-	SB	19-524	#1
3600	124.2	124.4	-	-	SB		
3601	115.9	116.2	116.2	-	SB		
3602	121.0	121.2	-	-	SB		
3603	115.9	116.0	-	-	SB		
3604	115.5	115.9	115.9	-	SB		
3605	124.3	124.5	-	-	SB	19-524	#2
3606	121.7	121.7	-	-	SB		
3607	115.9	116.0	-	-	SB		
3608	116.2	116.1	-	-	SB		
3609	123.1	122.9	-	-	SB		
3610	122.0	121.7	121.8	-	SB		
3611	115.6	115.6	-	-	SB	19-524	#3
3612	116.3	116.2	-	-	SB		

Weight 1 Date/Time:

7/24-8:00

Weight 2 Date/Time:

7/25-7:00

Weight 3 Date/Time:

7/26-7:45

Weight 4 Date/Time:

ASTM E2515 - Probes

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
1A	115630.4	115630.6	-	-	SB	19-509	#6
1B	115908.2	115904.4	-	-	SB		
2A	116241.8	116241.7	-	-	SB	19-509	#
2B	116331.8	116331.6	-	-	SB		
3A	116076.7	116076.8	-	-	SB	19-510	#1
3B	116341.4	116341.4	-	-	SB		
4A	116184.3	116184.2	-	-	SB	19-502	1
4B	116366.9	116366.9	-	-	SB		
5A	116769.5	116769.4	-	-	SB	19-502	2
5B	116877.5	116877.3	-	-	SB		

Weight 1 Date/Time:
7/22 - 7:00

Weight 2 Date/Time:
7/23 - 8:00

Weight 3 Date/Time:

Weight 4 Date/Time:

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
6A	116545.1	116545.2	-	-	L	19-502	3
6B	116118.4	116118.3	-	-	L		
7A	116741.2	116741.2	-	-	L	19-502	4
7B	117289.3	117289.1	-	-	L		
8A	116824.4	116824.4	-	-	L	19-501	1
8B	116826.3	116826.2	-	-	L		
9A	116714.1	116714.2	-	-	L	19-501	1
9B	117918.9	117918.8	-	-	L		
10A	-	116821.1	116821.3	-	L	19-494	#1
10B	-	117904.7	117904.9	-	L	19-494	#1

Weight 1 Date/Time:
8/15 0800

Weight 2 Date/Time:
8/16 0930

Weight 3 Date/Time:
8/5 15:45

Weight 4 Date/Time:

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
11A	117636.9	117636.7	-	-	SB	19-494	#2
11B	117490.9	117490.5	117490.7	-	SB	19-494	#2
12A	116889.6	116889.4	-	-	SB	19-494	#3
12B	117957.6	117957.5	-	-	SB	19-494	#3
13A	117456.4	117456.3	-	-	SB	19-494	#4
13B	117055.1	117055.2	-	-	SB	19-494	#4
14A	116818.4	-	116818.2	-	SB	19-494	#5
14B	116772.2	-	116772.0	-	SB	19-494	#5
15A	117419.0	-	-	-	-	-	-
15B	116905.8	-	-	-	-	-	-

Weight 1 Date/Time:
9/5/ 15:45

Weight 2 Date/Time:
9/9 8:00

Weight 3 Date/Time:
9/10 - 1400

Weight 4 Date/Time:

ASTM E2515 - O-Rings

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
1A	3565.5	3565.4	-	-	SB	19-509	#6
1B	3553.9	3554.1	-	-	SB		
2A	3551.2	3551.4	-	-	SB	19-509	#7
2B	3569.8	3570.0	-	-	SB		
3A	3578.3	3578.5	-	-	SB	19-510	#1
3B	3566.7	3567.3 → 3566.8	-	-	SB		
4A	3591.2	3591.3	-	-	SB	19-502	1
4B	3578.6	3578.8	-	-	SB		
5A	3532.9	3533.1	-	-	SB	19-502	2
5B	3529.3	3529.5	-	-	SB		

Weight 1 Date/Time:
7/27-7/15

Weight 2 Date/Time:
7/24-8:00

Weight 3 Date/Time:

Weight 4 Date/Time:

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
6A	3614.9	3614.7	-	-	A	19-502	3
6B	3395.9	3396.1	-	-	A		
7A	3572.6	3572.7	-	-	A	19-502	4
7B	3521.5	3521.7	-	-	A		
8A	3551.6	3551.5	-	-	A	19-511	1
8B	3585.3	3585.3	-	-	A		
9A	3581.0	3581.2	-	-	A	19-501	#1
9B	3524.0	3524.1	-	-	A		
10A	3430.9	3431.0	-	-	A	19-494	#1
10B	3570.0	3570.2	-	-	A		

Weight 1 Date/Time:
8/15 0800

Weight 2 Date/Time:
8/16 0930

Weight 3 Date/Time:

Weight 4 Date/Time:

Sample	Weight 1	Weight 2	Weight 3	Weight 4	Initial	Project	Run
11A	3424.3	3424.4	-	-	SB	19-494	#2
11B	4234.1	4234.2	-	-	SB		
12A	3396.0	3396.2	-	-	SB	19-494	#3
12B	3406.4	3406.6	-	-	SB		
13A	3361.1	3361.3	-	-	SB	19-494	#4
13B	3446.2	3446.1	-	-	SB		
14A	3367.2	3367.4	-	-	SB	19-494	#5
14B	3341.4	3341.6	-	-	SB		
15A	3569.9						
15B	3570.7						

Weight 1 Date/Time:
9/5/10 16:00

Weight 2 Date/Time:
9/9 8:00

Weight 3 Date/Time:

Weight 4 Date/Time:



QUALITY CONTROL SERVICES

LABORATORY EQUIPMENT • SALES • SERVICE • CALIBRATION • REPAIRS
2340 SE 11TH Ave. Portland, Oregon 97214 • Box 14831 Portland, Oregon 97293
(503) 236-2712 • FAX (503) 235-2535 • www.qc-services.com



PFS Teco
11785 SE Hwy 212 STE#305
Clackamas, OR 97015

Report Number: DIR101A05026181218

A2LA ACCREDITED **CERTIFICATE OF CALIBRATION WITH DATA**

INSTRUMENT INFORMATION

Item	Make	Model	Serial Number	Customer ID	Location
Scale	Rice Lake	IQ+355E-2A x 100l	A05026	#041	Lab
Units	Readability	SOP	Cal Date	Last Cal Date	Cal Due Date
lbs	0.1	QC033	12/18/18	6/13/18	12/2019

FUNCTIONAL CHECKS

SHIFT TEST	LINEARITY	REPEATABILITY	ENVIRONMENTAL CONDITIONS
Test Wt: 250 Tol: 1	Test Wt: HB44 Tol: HB44	Test Wt: 100 Tol: 1	
As-Found: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	As-Found: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	As-Found: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	<input type="checkbox"/> Good <input checked="" type="checkbox"/> Fair <input type="checkbox"/> Poor
As-Left: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	As-Left: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	As-Left: Pass: <input checked="" type="checkbox"/> Fail: <input type="checkbox"/>	Temperature: 16.9°C

CALIBRATION DATA

Standard	As-Found	As-Left	Expanded Uncertainty
1000	999.3	1000.2	0.12
700	699.7	700.1	0.12
500	499.7	500.1	0.08
300	299.8	300.1	0.08
100	99.9	100.0	0.05
50	50.0	50.0	0.05

CALIBRATION STANDARDS

Item	Make	Model	Serial Number	Cal Date	Cal Due Date	NIST ID
Avoirdupois Cast W	Rice Lake	25 and 50lb	PWO990-CA	11/24/17	11/2019	20172265

Permanent Information Concerning this Equipment:

12 month calibration cycle. 2000lb platform.

Comments/Information Concerning this Calibration

12/18 - RH = 67%. Adjusted span.

Report prepared/reviewed by: ServiceTech Date: 12/28/18

Technician: R. Kauble

Signature: [Signature]

THIS CERTIFICATE SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE APPROVAL OF QUALITY CONTROL SERVICES, INC.

The uncertainty is calculated according to the ISO Guide to the Expression of Uncertainty in Measurement and includes the uncertainty of standards used combined with the observed standard deviation of the unit under test. The uncertainty is expanded with a k factor of 2 for an approximate 95% level of confidence. Instruments listed above were calibrated using standards traceable to the National Institute of Standards and Technology (NIST). Calibration data reflect results at the time and location of calibration. Calibration data should be reviewed to insure that the instrument is performing to its required accuracy.

Dry Gas Meter Calibration

Meter Manufacturer: Apex
 Model: XC-60-ED
 Lab ID #: 53
 Serial #: 1902130
 Calibration Date: 6/14/2019
 Calibration Expiration: 12/14/2019
 Barometric Pressure: 29.97 in. Hg



Reference Standard DGM	
Manufacturer:	Apex
Model:	SK25DA
Lab ID#:	47
Serial #:	1101001
Calibration Expiration Date:	3/13/2020
Calibration γ Factor:	0.998

Unit Under Test Previous Calibration	
Date	12/17/2018
γ Factor:	1.004
Allowable Deviation ($\pm 5\%$):	0.0502
Actual Deviation:	0.01
Result:	PASS

Calibration Data	Run 1	Run 2	Run 3
Standard DGM Initial Volume (L)	0.000	0.000	0.000
Standard DGM Final Volume (L)	147.373	142.005	143.359
Standard DGM Temperature ($^{\circ}\text{F}$)	71.0	72.0	72.0
Standard DGM Pressure (in H_2O)	0.00	0.00	0.0
DGM Initial Volume (ft^3)	0.000	0.000	0.000
DGM Final Volume (ft^3)	5.227	5.114	5.224
DGM Temperature ($^{\circ}\text{F}$)	78.0	85.0	91.0
DGM Pressure (in H_2O)	2.67	2.00	1.5
Time (min)	33.0	36.0	42.0
Net Volume for Standard DGM (ft^3)	5.204	5.015	5.063
Net Volume for DGM (ft^3)	5.227	5.114	5.224

Dry Gas Meter γ Factor	1.000	0.998	0.998
γ Factor Deviation From Average	1.000	0.998	0.998

Average Gas Meter γ Factor

0.999

Calculations:

- Deviation = |Average value for all runs - current run value|
- $\gamma = [V_{\text{std}} \times (\gamma_{\text{std}}) \times (P_{\text{bar}} + P_{\text{std}}/13.6) \times (T_{\text{DGM}} + 460)] / [V_{\text{DGM}} \times (T_{\text{std}} + 460) \times (P_{\text{bar}} + P_{\text{DGM}}/13.6)]$

Standard Reference Meter is calibrated to NIST traceable standards. Uncertainty of measurement is $\pm 0.5\%$.

Dry Gas Meter Calibration

Meter Manufacturer: Apex

Model: XC-60-ED

Lab ID #: 54

Serial #: 1902133

Calibration Date: 6/14/2019

Calibration Expiration: 12/14/2019

Barometric Pressure: 29.97 in. Hg



Reference Standard DGM

Manufacturer:	Apex
Model:	SK25DA
Lab ID#:	47
Serial #:	1101001
Calibration Expiration Date:	3/13/2020
Calibration γ Factor:	0.998

Unit Under Test Previous Calibration

Date	12/17/2018
γ Factor:	1.000
Allowable Deviation ($\pm 5\%$):	0.05
Actual Deviation:	0.00
Result:	PASS

Calibration Data

	Run 1	Run 2	Run 3
Standard DGM Initial Volume (L)	0.000	0.000	0.000
Standard DGM Final Volume (L)	139.967	143.359	139.656
Standard DGM Temperature ($^{\circ}$ F)	72.0	73.0	75.0
Standard DGM Pressure (in H ₂ O)	0.00	0.00	0.0
DGM Initial Volume (ft ³)	0.000	0.000	0.000
DGM Final Volume (ft ³)	5.098	5.242	5.114
DGM Temperature ($^{\circ}$ F)	92.0	93.0	95.0
DGM Pressure (in H ₂ O)	2.99	2.02	1.3
Time (min)	30.0	37.0	45.0
Net Volume for Standard DGM (ft ³)	4.943	5.063	4.932
Net Volume for DGM (ft ³)	5.098	5.242	5.114

Dry Gas Meter γ Factor	0.997	0.995	0.995
γ Factor Deviation From Average	0.997	0.995	0.995

Average Gas Meter γ Factor

0.996

Calculations:

1. Deviation = |Average value for all runs - current run value|

2. $\gamma = [V_{std} \times (\gamma_{std}) \times (P_{bar} + P_{std}/13.6) \times (T_{DGM} + 460)] / [V_{DGM} \times (T_{std} + 460) \times (P_{bar} + P_{DGM}/13.6)]$

Standard Reference Meter is calibrated to NIST traceable standards. Uncertainty of measurement is $\pm 0.5\%$.



QUALITY CONTROL SERVICES

LABORATORY EQUIPMENT • SALES • SERVICE • CALIBRATION • REPAIRS
2340 SE 11TH Ave. Portland, Oregon 97214 • Box 14831 Portland, Oregon 97293
(503) 236-2712 • FAX (503) 235-2535 • www.qc-services.com



Report of Calibration

Firm: Dirigo Laboratories
Address: 11785 SE Hwy 212, Ste 305
City/State/Zip: Clackamas, OR 97015

Test Completed: 03/21/17
Submitted By: John Steiner
Traceable Number: 20170468

Test Item: 200mg and 100mg Individual Weights
Serial No.: Listed in Table

Manufacturer: Troemner

<u>Material</u>	<u>Assumed Density</u>	<u>Range</u>	<u>Tolerance Class</u>
Stainless Steel	7.95 g/cm ³	200mg & 100mg	ASTM Class 1

Method and Traceability

The procedure used for this calibration is NIST IR 6969 SOP 4 Double Substitution Weighing Design. Standards used for comparison are traceable to the National Institute of Standards and Technology (reports on file) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and traceability within the level of uncertainty reported. The Traceable Number listed above is Traceable to National Standards through an unbroken chain of comparison each having stated uncertainties.

Standards Used:

100g to 1mg Working Standards Were Calibrated: 03/03/17 Due: 03/31/18 Standards ID: 723318

Mass Comparators Used: MET-05

Tested by: D. Thompson

Conventional Mass: "The conventional value of the result of weighing a body in air is equal to the mass of a standard, of conventionally chosen density, at a conventionally chosen temperature, which balances this body at this reference temperature in air of conventionally chosen density. International Recommendation 33 (OIML IR 33 1973, 1979). "Conventional Value of the Result of Weighing in Air" (Previously known as "Apparent Mass vs. 8.0g/cm³).

Uncertainty Statement: The uncertainty conforms to the ISO Guide to the Expressions of Uncertainty in Measurement. Uncertainty as reported is based on a coverage factor k=2 for an approximate 95 percent level of uncertainty. Uncertainty components include the standard deviation of the process, the uncertainty of the standard used, an uncertainty component associated with the potential drift of the standard used, and the estimated uncertainty related to measuring and determining the air buoyancy effect.

Conventional Mass Values are listed on page 2 of this report.

page 1 of 2

Quality Control Services, Inc.
Metrology Laboratory Manager
E-mail dthompson@qc-services.com

Date: 03/21/17

Signature David S. Thompson

This document shall not be reproduced, except in full, without the written approval of Quality Control Services Mass Laboratory.

Member: National Conference of Standards Laboratories and Weights & Measures



QUALITY CONTROL SERVICES

LABORATORY EQUIPMENT • SALES • SERVICE • CALIBRATION • REPAIRS
2340 SE 11TH Ave. Portland, Oregon 97214 • Box 14831 Portland, Oregon 97293
(503) 236-2712 • FAX (503) 235-2535 • www.qc-services.com



Report of Calibration

Firm: Dirigo Laboratories
Address: 11785 SE Hwy 212, Ste 305
City/State/Zip: Clackamas, OR 97015

Test Completed: 03/21/17
Submitted By: John Steiner
Traceable Number: 20170468

Test Item: 200mg and 100mg Individual Weights
Serial No.: Listed in Table

Manufacturer: Troemner

Laboratory Environment at time of test

Temperature °C	Pressure mmHg	Humidity %RH
21.967	753.44	49.44

Conventional Mass Value

Nominal Value	As Found grams	As Found Correction* (mg)	Uncertainty (mg)	Tolerance (mg)
200mg SN 1000101395	0.2000061	0.0061	0.0026	0.01
100mg SN 1000126267	0.1000046	0.0046	0.0028	0.01

*Correction is the difference between the conventional mass value of a weight and its nominal value.

Comments: These weights were new from the manufacturer and were within ASTM Class 1 tolerances As Found. No adjustments or changes were made so As Found values should be considered to be As Left values.

Accredited by the American Association for Laboratory Accreditation (A2LA) under Calibration Laboratory Code 115953 and Certificate Number 1550.01. This laboratory meets the requirements of ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration.

page 2 of 2

Quality Control Services, Inc.
Metrology Laboratory Manager
E-mail dthompson@qc-services.com

Date: 03/21/17

Signature David S. Thompson



QUALITY CONTROL SERVICES

LABORATORY EQUIPMENT • SALES • SERVICE • CALIBRATION • REPAIRS
2340 SE 11TH Ave. Portland, Oregon 97214 • Box 14831 Portland, Oregon 97293
(503) 236-2712 • FAX (503) 235-2535 • www.qc-services.com



Report of Calibration

Firm: Dirigo Laboratories
Address: 11785 SE Hwy 212, Ste 305
City/State/Zip: Clackamas, OR 97015

Test Completed: 01/15/16
Purchase Order: 1001
Traceable Number: 20152489

Test Item: 20lb and 10lb Individual Grip Handle Weights
Serial No.: Listed in Table

Manufacturer: Unknown

Laboratory Environment at time of test

Temperature °C	Pressure mmHg	Humidity %RH
21.448	760.64	44.58

Conventional Mass Value

Nominal Value	As Found pounds	As Found Correction* (mg)	Uncertainty (mg)	Tolerance (mg)
20lb #098	19.9995450	-206.4	6.4	910
10lb #097	10.0006510	295.3	5.1	450
10lb #051	10.0003421	155.2	5.1	450

*Correction is the difference between the conventional mass value of a weight and its nominal value.


Comments: These weights were received in good condition and were within NIST Handbook 105-1 Class F tolerances As Found. No adjustments or changes were made so As Found values should be considered to be As Left values.

Accredited by the American Association for Laboratory Accreditation (A2LA) under Calibration Laboratory Code 115953 and Certificate Number 1550.01. This laboratory meets the requirements of ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and any additional program requirements in the field of calibration.

page 2 of 2

Quality Control Services, Inc.
Metrology Laboratory Manager
E-mail dthompson@qc-services.com

Date: 01/15/16


Signature David S. Thompson



Established 1974

QUALITY CONTROL SERVICES

LABORATORY EQUIPMENT • SALES • SERVICE • CALIBRATION • REPAIRS
2340 SE 11TH Ave. Portland, Oregon 97214 • Box 14831 Portland, Oregon 97293
(503) 236-2712 • FAX (503) 235-2535 • www.qc-services.com



PFS Teco
11785 SE Hwy 212 STE#305
Clackamas, OR 97015

Report Number: DIRI0134307497181218

A2LA ACCREDITED CERTIFICATE OF CALIBRATION WITH DATA

INSTRUMENT INFORMATION

Item	Make	Model	Serial Number	Customer ID	Location
Balance	Sartorius	ENTRIS224-1S	34307497	#107	Lab
Units	Readability	SOP	Cal Date	Last Cal Date	Cal Due Date
g	0.0001	QC012	12/18/18	6/13/18	12/2019

FUNCTIONAL CHECKS

ECCENTRICITY		LINEARITY		STANDARD DEVIATION			ENVIRONMENTAL CONDITIONS		
Test Wt:	Tol:	Test Wt:	Tol:	Test Wt:	Tol:		<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
100	0.0003	50 x 4	0.0002	100	0.0001		Good	Fair	Poor
As-Found:		As-Found:		1. 100.0001	5. 100.0002	9. 100.0001	Temperature: 21.3°C		
Pass: <input checked="" type="checkbox"/>	Fail: <input type="checkbox"/>	Pass: <input checked="" type="checkbox"/>	Fail: <input type="checkbox"/>	2. 100.0001	6. 100.0001	10. 100.0001			
As-Left:		As-Left:		3. 100.0001	7. 100.0001	Result			
Pass: <input checked="" type="checkbox"/>	Fail: <input type="checkbox"/>	Pass: <input checked="" type="checkbox"/>	Fail: <input type="checkbox"/>	4. 100.0001	8. 100.0002	0.00004			

A2LA ACCREDITED SECTION OF REPORT

Standard	As-Found	As-Left	Expanded Uncertainty
200	200.0002	200.0001	0.00014
100	100.0001	100.0001	0.00014
50	50.0003	50.0001	0.00014
20	20.0001	20.0001	0.00014
1	1.0001	1.0000	0.00014
0.1	0.1000	0.1000	0.00014

CALIBRATION STANDARDS

Item	Make	Model	Serial Number	Cal Date	Cal Due Date	NIST ID
Weight Set	R.L./Troemner	10kg to 1mg	G782	1/3/18	1/2019	20172421

Permanent Information Concerning this Equipment:

12 month calibration cycle.

Comments/Info Concerning this Calibration:

12/18 - RH = 56%. Adjusted span.

Report prepared/reviewed by: ServiceTech X Date: 12/28/18

Technician: R. Kauble

Signature:

THIS CERTIFICATE SHALL NOT BE REPRODUCED WITHOUT THE APPROVAL OF QUALITY CONTROL SERVICES, INC.

The uncertainty is calculated according to the ISO Guide to the Expression of Uncertainty in Measurement and includes the uncertainty of standards used combined with the observed standard deviation and readability of the unit under test. The uncertainty is expanded with a k factor of 2 for an approximate 95% level of confidence. Instruments listed above were calibrated using standards traceable to the National Institute of Standards and Technology (NIST). Calibration data reflect results at the time and location of calibration. Calibration data should be reviewed to insure that the instrument is performing to its required accuracy. Calibrations comply with ISO/IEC 17025 and ANSI/Z540-1-1994 quality standards.

Member: National Conference of Standards Laboratories and Weights & Measures

PT ID: DIRI01

CERTIFICATE OF CALIBRATION

CUSTOMER: PFS-TECO : CLACKAMAS, OR
PO NUMBER: N/A
INST. MANUFACTURER: DWYER
INST. DESCRIPTION: VELOMETER
MODEL NUMBER: 471
SERIAL NUMBER: CP288559 (ID# 095)
RATED UNCERTAINTY: SEE NOTES BELOW.
UNCERTAINTY GIVEN: $\pm .20\%$ RD ; k=2

CALIBRATION DATE: 03/14/2019
CALIBRATION DUE: 03/14/2020
PROCEDURE: T.O.33K6-4-1769-1
CALIBRATION FLUID: AIR @ 14.7 PSIA 70°F
RECEIVED CONDITION: WITHIN MFG. SPECS.
LEFT CONDITION: WITHIN MFG. SPECS.
AMBIENT CONDITIONS: 762 mm HGA 43% RH 69°F
CERTIFICATE FILE #: 490265.2019

NOTES: $\pm 3\%$ FS (0-500 / 0-1500) *** $\pm 4\%$ F.S. (0-5000) *** $\pm 5\%$ F.S. (0-15000) *** $\pm 2^\circ\text{F}$
NOTES CONT. : Q.MANUAL IM 1.5 REV 2017.1 DATED 7-18-2017

UUT INDICATED FT/MIN	DM.STD. ACTUAL FT/MIN	UUT INDICATED DEG. F	DM STD. ACTUAL DEG. F
64	65	0 TO 200°F	0 TO 200°F
110	112	43.4	43.5
206	210	69.0	68.9
498	509	99.4	99.2
503	505		
1049	1058		
1497	1514		
509	513		
3419	3460		
4992	5068		
5136	5235		
13928	14232		

STANDARDS USED:

A220: 12" WIND TUNNEL 0 - 8000 FPM CMC $\pm .203\%$ RD TRACE# 1520423238	DUE	05/23/2019
A24: HART SCIENTIFIC TEMP. STANDARD $\pm .024^\circ\text{F}$ TRACE# 1520423238	DUE	03/07/2020

All instruments used in the performance of the shown calibration have traceability to the National Institute of Standards and Technology (NIST). The uncertainty ratio between the calibration standards (DM.STD.) used and the unit under test (UUT) is a minimum of 4:1, unless otherwise noted. Calibration has been performed per the shown procedure number, in accordance with ISO 10012:2003, ISO 17025:2005, ANSI/NC SL-Z-540.3, and/or MIL-STD-45662A. Test methods: API2530-92 & ASME MFC-3M-1989.

Dick Munns Company • 11133 Winners Circle • Los Alamitos, CA 90720
Phone (714) 827-1215 • Fax (714) 827-0823

This Calibration Certificate shall not be reproduced, except in full, without approval by DICK MUNN'S COMPANY. The data shown applies only to the instrument being calibrated and under the stated conditions of calibration.

Date:

3/14/2019

Approved By:

[Signature]

Calibration Technician:

[Signature]



Model 1430 Microtector® Electronic Point Gage

Installation and Operating Instructions



Model 1430 Microtector® Portable Electronic Point Gage combines modern, solid-state integrated circuit electronics with a time-proven point gage manometer to provide fast, accurate pressure measurements.

SPECIFICATIONS AND FEATURES

- Accurate and repeatable to $\pm .00025$ inches water column
- Pressure range: 0 - 2" w.c., positive, negative, or differential pressures
- Non-toxic and inexpensive gage fluid consists of distilled water mixed with a small amount of fluorescein green color concentrate
- Convenient, portable, lightweight and self-contained, the unit requires no external power connections and is operated by a 1.5 volt penlight cell
- A.C. detector current eliminates point plating, fouling and erosion
- Micrometers are manufactured in accordance with ASME B89.1.13-2001, and are traceable to a standard at the National Institute of Standards and Technology
- Three-point mounting, dual leveling adjustment, and circular level vial assure rapid setup
- Durablock® precision-machined acrylic plastic gage body
- Sensitive 0 - 50 microamp D.C. meter acts as a detector and also indicates battery and probe condition
- Heavy 2" thick steel base plate provides steady mounting
- Top-quality glass epoxy circuit board and solid-state, integrated circuit electronics
- Electronic enclosure of tough, molded styrene acrylonitrile provides maximum protection to components yet allows easy access to battery compartment
- Rugged sheet steel cover and carrying case protects the entire unit when not in use
- Accessories included are (2) 3-foot lengths Tygon® tubing, (2) 1/8" pipe thread adapters and 3/4 oz. bottle of fluorescein green color concentrate with wetting agent

Maximum pressure: 100 psig with optional pipe thread connections.

Tygon® is a registered trademark of Saint-Gobain Corporation

DWYER INSTRUMENTS, INC.

P.O. BOX 373

MICHIGAN CITY, INDIANA 46361, U.S.A

Phone: 219/879-8000

Fax: 219/872-9057

www.dwyer-inst.com

e-mail: info@dwyer-inst.com



CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PXPKG TUALATIN OR H
10450 SW TUALATIN SHERWOOD ROAD
TUALATIN OR 97062

Certificate Modification Date: 10/01/2018

Praxair Order Number: 70743165

Part Number: NI CD17CO8E-AS

Fill Date: 09/26/2018

Lot Number: 70086826911

Cylinder Style & Outlet: AS

CGA 590

Cylinder Pressure and Volume: 1290 psig 140 ft3

Certified Concentration

Expiration Date:	10/01/2026	NIST Traceable
Cylinder Number:	SA17187	Expanded Uncertainty
17.00 %	Carbon dioxide	± 0.3 %
4.31 %	Carbon monoxide	± 0.6 %
16.99 %	Oxygen	± 0.2 %
Balance	Nitrogen	

ProSpec EZ Cert



Certification Information:

Certification Date: 10/01/2018

Term: 96 Months

Expiration Date: 10/01/2026

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.

Do Not Use this Standard if Pressure is less than 100 PSIG.

CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candi:late)

1. Component:

Carbon dioxide

Requested Concentration: 17 %
Certified Concentration: 17.00 %
Instrument Used: Horiba VIA-510 S/N 20C194WK
Analytical Method: NDIR
Last Multipoint Calibration: 09/21/2018

First Analysis Data:				Date
Z:	0	R:	20.1	10/01/2018
R:	20.1	Z:	0	
Z:	0	C:	17.01	
C:	17	R:	20.11	
Conc:	17	Conc:	17.01	
UOM:	%	Mean Test Assay:	17	%

Reference Standard:

Type / Cylinder #: GMIS / CC187238

Concentration / Uncertainty: 20.10 % ±0.24%

Expiration Date: 06/07/2026

Traceable to: SRM # / Sample # / Cylinder #: RGM#CC193512 / N/A / RGM#CC193512

SRM Concentration / Uncertainty: 26.99% / ±0.05%

SRM Expiration Date: 05/15/2023

Second Analysis Data:				Date
Z:	0	R:	0	
R:	0	Z:	0	
Z:	0	C:	0	
C:	0	R:	0	
Conc:	0	Conc:	0	
UOM:	%	Mean Test Assay:		%

2. Component:

Carbon monoxide

Requested Concentration: 4.25 %
Certified Concentration: 4.31 %
Instrument Used: Horiba VIA-510 S/N UB9UCSYX
Analytical Method: NDIR
Last Multipoint Calibration: 09/21/2018

First Analysis Data:				Date
Z:	0	R:	5	10/01/2018
R:	5	Z:	0	
Z:	0	C:	4.32	
C:	4.31	R:	5.01	
Conc:	4.31	Conc:	4.3	
UOM:	%	Mean Test Assay:	4.31	%

Reference Standard:

Type / Cylinder #: GMIS / CC242633

Concentration / Uncertainty: 5.00 % ±0.543%

Expiration Date: 04/03/2025

Traceable to: SRM # / Sample # / Cylinder #: SRM 2642a / 51-D-23 / FF23106

SRM Concentration / Uncertainty: 7.859% / ±0.039%

SRM Expiration Date: 07/15/2019

Second Analysis Data:				Date
Z:	0	R:	0	
R:	0	Z:	0	
Z:	0	C:	0	
C:	0	R:	0	
Conc:	0	Conc:	0	
UOM:	%	Mean Test Assay:		%

3. Component:

Oxygen

Requested Concentration: 17 %
Certified Concentration: 16.99 %
Instrument Used: OXYMAT 5E
Analytical Method: Paramagnetic
Last Multipoint Calibration: 09/04/2018

First Analysis Data:				Date
Z:	0	R:	20.86	10/01/2018
R:	20.86	Z:	0	
Z:	0	C:	16.99	
C:	16.99	R:	20.86	
Conc:	16.99	Conc:	16.99	
UOM:	%	Mean Test Assay:	16.99	%

Reference Standard:

Type / Cylinder #: GMIS / CC75874

Concentration / Uncertainty: 20.86 % ±0.111%

Expiration Date: 11/07/2025

Traceable to: SRM # / Sample # / Cylinder #: SRM 2659a / 71-E-19 / FF22331

SRM Concentration / Uncertainty: 20.863% / ±0.021%

SRM Expiration Date: 08/23/2021

Second Analysis Data:				Date
Z:	0	R:	0	
R:	0	Z:	0	
Z:	0	C:	0	
C:	0	R:	0	
Conc:	0	Conc:	0	
UOM:	%	Mean Test Assay:		%

Analyzed By

Jose Vasquez

Certified By

Danielle Burns



CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PXPKG TUALATIN OR H
10450 SW TUALATIN SHERWOOD ROAD
TUALATIN OR 97062

Certificate Modification Date: 09/05/2018

Praxair Order Number: 70716136

Part Number: NI CD10CO33E-AS

Fill Date: 08/31/2018

Lot Number: 70086824308

Cylinder Style & Outlet: AS CGA 590
Cylinder Pressure and Volume: 2000 psig 140 ft3

Certified Concentration

Expiration Date:	09/05/2026	NIST Traceable
Cylinder Number:	CC170624	Expanded Uncertainty
10.00 %	Carbon dioxide	± 0.3 %
2.51 %	Carbon monoxide	± 0.7 %
10.50 %	Oxygen	± 0.6 %
Balance	Nitrogen	

ProSpec EZ Cert



Certification Information:

Certification Date: 09/05/2018

Term: 96 Months

Expiration Date: 09/05/2026

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard if Pressure is less than 100 PSIG.

CO responses have been corrected for CO2 interference. CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Carbon dioxide

Requested Concentration: 10 %
Certified Concentration: 10.00 %
Instrument Used: Horiba VIA-510 S/N 20C194WK
Analytical Method: NDIR
Last Multipoint Calibration: 08/20/2018

First Analysis Data:				Date
Z: 0	R: 14.02	C: 10	Conc: 10	09/05/2018
R: 14.02	Z: 0	C: 10	Conc: 10	
Z: 0	C: 10	R: 14.02	Conc: 10	
UOM: %				
Mean Test Assay:				10 %

Reference Standard: Type / Cylinder #: GMIS / CC141375

Concentration / Uncertainty: 14.02 % ± 0.3%

Expiration Date: 06/11/2026

Traceable to: SRM # / Sample # / Cylinder #: SRM 1675b / 6-F-51 / CAL014538

SRM Concentration / Uncertainty: 13.963% / ± 0.034%

SRM Expiration Date: 05/16/2022

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: %				
Mean Test Assay:				%

2. Component: Carbon monoxide

Requested Concentration: 2.5 %
Certified Concentration: 2.51 %
Instrument Used: Horiba VIA-510 S/N UB9UCSYX
Analytical Method: NDIR
Last Multipoint Calibration: 08/20/2018

First Analysis Data:				Date
Z: 0	R: 2.48	C: 2.51	Conc: 2.51	09/05/2018
R: 2.48	Z: 0	C: 2.51	Conc: 2.51	
Z: 0	C: 2.51	R: 2.48	Conc: 2.51	
UOM: %				
Mean Test Assay:				2.51 %

Reference Standard: Type / Cylinder #: GMIS / CC102045

Concentration / Uncertainty: 2.48 % ± 0.448%

Expiration Date: 04/03/2025

Traceable to: SRM # / Sample # / Cylinder #: SRM 2641a / 52-D-30 / CAL017193

SRM Concentration / Uncertainty: 4.009% / ± 0.017%

SRM Expiration Date: 07/15/2019

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: %				
Mean Test Assay:				%

3. Component: Oxygen

Requested Concentration: 10.5 %
Certified Concentration: 10.50 %
Instrument Used: OXYMAT 5E
Analytical Method: Paramagnetic
Last Multipoint Calibration: 09/04/2018

First Analysis Data:				Date
Z: 0	R: 9.88	C: 10.49	Conc: 10.49	09/05/2018
R: 9.88	Z: 0	C: 10.5	Conc: 10.5	
Z: 0	C: 10.5	R: 9.88	Conc: 10.5	
UOM: %				
Mean Test Assay:				10.5 %

Reference Standard: Type / Cylinder #: NTRM / DT0010402

Concentration / Uncertainty: 9.88 % ± 0.4%

Expiration Date: 11/18/2022

Traceable to: SRM # / Sample # / Cylinder #: NTRM #170701 / NIA / NTRM #DT0010402

SRM Concentration / Uncertainty: 9.875% / ± 0.040%

SRM Expiration Date: 11/18/2022

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: %				
Mean Test Assay:				%

Analyzed By

Danielle Burns

Certified By

José Vasquez

Verification of Standardization

of Tape Measure

by
Advanced Calibration Technologies
28111 S.E. Wally Road
Boring, OR 97009
1-800-259-5058



Customer:	PFS Teco, Inc	Street:	11785 Southeast Highway 212 Suite 305
City:	Clackamas	State:	OR
Machine Manufacturer:	Dewalt	Zip:	97015 Location: In House
Capacity:	0.000 - 192.000 inches 0.125 Divisions	Model:	16' Tape Measure
Calibration Cycle:	12 Months	Serial #:	090
Previous Calibration Date:	January 2019	Lab ID#:	#090
Equipment Used:	Gauge Blocks S/N: ADGB002	Calibration Procedure:	Ad-Tek SR
If Other, Explain:		Action Recommended:	

Verification Data

Purpose: This method provides instructions for checking the critical dimensions of the equipment.			
Tolerance: Equipment shall meet the dimensional tolerances specified in the applicable test method.			
Procedure: Verified using manufacturer's procedures.			
Actual Dimensions (inches)	Unit Under Test As Found (inches)	Unit Under Test As Left (inches)	Difference (inches)
0.0000	0.000	0.000	0.000
0.1250	0.050	0.050	-0.075
0.2500	0.250	0.250	0.000
0.5000	0.500	0.500	0.000
0.7500	0.750	0.750	0.000
1.0000	1.000	1.000	0.000
3.0000	3.000	3.000	0.000
5.0000	5.000	5.000	0.000
7.0000	7.000	7.000	0.000
9.0000	9.000	9.000	0.000
12.0000	12.000	12.000	0.000
The overall condition of the device as found:		Within Specification	
The overall condition of the device as left:		Within Specification	
The measurement of uncertainty (MU) was calculated to be:		0.00060	

File No: PFS-101666-0119D0120-AH-SR-090

Temperature: 72.1°F Humidity: 41.1%

The equipment used in the verification of this instrument has been calibrated and is NIST traceable.
The uncertainty of calibration was estimated at the 95% confidence level, coverage factor (k=2).

Remarks:

This certificate of verification is issued as a statement of fact that on the date of verification the above instrument had an accuracy as indicated and was calibrated to meet the requirements of the manufacturer's specifications. This certificate should not be construed or regarded as a guarantee or warranty of any kind that the instrument will retain the same percentage of accuracy as determined on the date when the verification was performed and reported. Ad-Tek, Inc. hereby expressly disclaims any and all liability for damage or loss by all parties arising or resulting from deterioration, obsolescence, malfunction, subsequent calibration performed by another agency or substandard performance of said instrument.

This report and certificate of verification shall not be reproduced except in full, without the written approval of Ad-Tek, Inc.


Service Technician:	<u>Alisa Houser</u>	Date of Service:	<u>January 16, 2019</u>
Technical Manager:	<u>Nicole Ostrowski</u>	Date Next Due:	<u>January 2020</u>

We sincerely appreciate your business and thank you for selecting Advanced Calibration Technologies, Inc. for servicing your equipment.
To reschedule, please call (800) 259-5058. Thank You.

Verification of Standardization

of Calipers

by
Advanced Calibration Technologies
28111 S.E. Wally Road
Boring, OR 97009
1-800-259-5058



Customer:	PFS Teco, Inc	Street:	11785 Southeast Highway 212 Suite 305
City:	Clackamas	State:	OR
Machine Manufacturer:	General	Zip:	97015
Capacity:	0.0000 - 6.0000 inches	0.0005 Divisions	Location: In House
Calibration Cycle:	12 Months	Model:	6" Digital Caliper
Previous Calibration Date:	January 2018	Serial #:	092
Equipment Used:	Gauge Blocks S/N: ADGB002	Lab ID#:	092
If Other, Explain:		Calibration Procedure:	Ad-Tek DC
		Action Recommended:	

Verification Data

Purpose: This method provides instructions for checking the critical dimensions of the inside diameter of the equipment.			
Tolerance: Equipment shall meet the dimensional tolerances specified by the manufacturer for the inside diameter.			
Procedure: Verified using the procedure to meet manufacturer's tolerance for inside diameter.			
Actual Dimensions (inches)	Unit Under Test As Found (inches)	Unit Under Test As Left (inches)	Difference (inches)
0.0000	0.0000	0.0000	0.0000
0.0500	0.0500	0.0500	0.0000
0.1000	0.1000	0.1000	0.0000
0.1010	0.1010	0.1010	0.0000
0.1050	0.1050	0.1050	0.0000
0.1100	0.1100	0.1100	0.0000
0.1500	0.1500	0.1500	0.0000
0.5000	0.5000	0.5000	0.0000
1.0000	1.0000	1.0000	0.0000
3.0000	2.9995	2.9995	-0.0005
5.0000	4.9990	4.9990	-0.0010
The overall condition of the device as found:		Within Specification	
The overall condition of the device as left:		Within Specification	
The measurement of uncertainty (MU) was calculated to be:		0.00062	

This certificate does not reflect measurements for inside jaws, step height, or depth.

File No: PFS-101666-0119D0120-AH-DC-092

Temperature: 68.2°F Humidity: 41.6%

The equipment used in the verification of this instrument has been calibrated and is NIST traceable.
The uncertainty of calibration was estimated at the 95% confidence level, coverage factor (k=2).

Remarks: _____

This certificate of verification is issued as a statement of fact that on the date of verification the above instrument had an accuracy as indicated and was calibrated to meet the requirements of the manufacturer's specifications. This certificate should not be construed or regarded as a guarantee or warranty of any kind that the instrument will retain the same percentage of accuracy as determined on the date when the verification was performed and reported. Ad-Tek, Inc. hereby expressly disclaims any and all liability for damage or loss by all parties arising or resulting from deterioration, obsolescence, malfunction, subsequent calibration performed by another agency or substandard performance of said instrument.

This report and certificate of verification shall not be reproduced except in full, without the written approval of Ad-Tek, Inc.

Service Technician: Alisa Houser Date of Service: January 15, 2019
Technical Manager: Nicole Ostrowski Date Next Due: January 2020

We sincerely appreciate your business and thank you for selecting Advanced Calibration Technologies, Inc. for servicing your equipment.
To reschedule, please call (800) 259-5058. Thank You.